

Brand HP
 Model 4195A
 Serial Nbr 2904J03383
 Date 3/31/17
 Temp
 Warm up 30 min



Equipment Instrument
 Rohde & Schwarz SMV03 Signal Gnerator
 Giga-tronics 8652A RF Power Meter

Reference Oscillator

	Units	Reading	PPM
10 MHz	MHz	10	-0.0697

Amplitude and Frequency Accuracy

Settings

Freq Span 10 MHz

Ref Osc External

Signal Applied	Units	Freq	Amplitude	Power	Me	Diff
100	MHz	100.09	-9.7	-10.43		0.73
200		200.09	-9.9	-10.58		0.68
400		400.09	-10.22	-10.87		0.65
800		800.09	-10.95	-11.39		0.44
1600		1300.09	-11.22	-12.05		0.83
2600		2600.09	-13.15	-13.09		-0.06

GPIB Control PASS