

**Silicon Valley Test & Meas.  
Performance Test Report**

9/19/18 21:03  
Stock Number: 513  
Serial Number: GB1080203

DUT Driver: DUT Driver.vi

Protocol: TP015-01



**DEVICE UNDER TEST**

Agilent Technologies,N3280A,GB41080203,A.00.01b

**INSTRUMENTS USED**

HEWLETT-PACKARD,34401A,0,11-5-2

**CH 1 PROGRAMMING AND MEASUREMENT ACCURACY**

PROG. VALUE	MEAS. VALUE	READBACK
0V	0.000128	-0.000139
10V	9.998467	9.997720
-10V	-9.997976	-9.998000
0.5 mA	0.000501	0.000500

**CH 2 PROGRAMMING AND MEASUREMENT ACCURACY**

PROG. VALUE	MEAS. VALUE	READBACK
0V	-0.000138	0.000025
10V	9.999276	9.998530
-10V	-9.999311	-9.999130
0.5 mA	0.000501	0.000500

**CH 3 PROGRAMMING AND MEASUREMENT ACCURACY**

PROG. VALUE	MEAS. VALUE	READBACK
0V	0.000229	-0.000765
10V	9.999505	9.997470
-10V	-9.998745	-9.999080
0.5 mA	0.000500	0.000500

**CH 4 PROGRAMMING AND MEASUREMENT ACCURACY**

PROG. VALUE	MEAS. VALUE	READBACK
0V	0.000167	-0.000479
10V	9.998349	9.997230
-10V	-9.997753	-9.998190
0.5 mA	0.000502	0.000500